

# CERTIFICATE OF COMPLIANCE

**Certificate Number** 20141124-E471172  
**Report Reference** E471172-20140930  
**Issue Date** 2014-NOVEMBER-24

**Issued to:** Flintec Inc  
18A Kane Industrial Dr  
Hudson MA 01749

**This is to certify that  
representative samples of**

**PROCESS CONTROL EQUIPMENT FOR USE IN  
HAZARDOUS LOCATIONS**

Intrinsically Safe for use in Class I, Division 1, Groups  
A, B, C, and D; Class II, Groups E, F, and G; and Class III. Hazardous  
Locations.

Models CC1 and CC1B, intrinsically safe load cells when installed in  
accordance with manufacturer's control drawing nos. 0061571 and  
0016640, respectively

Have been investigated by UL in accordance with the  
Standard(s) indicated on this Certificate.

**Standard(s) for Safety:**

UL 913 : Standard for Intrinsically Safe Apparatus and  
Associated Apparatus for Use in Class I, II, and III, Division  
1, Hazardous (Classified) Locations  
CAN/CSA C22.2 No. 157-92 : Standard for Intrinsically Safe  
and Non-incendive Equipment for Use in Hazardous  
Locations


**Additional Information:**

See the UL Online Certifications Directory at  
[www.ul.com/database](http://www.ul.com/database) for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's  
Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.

This is to certify that representative samples of the product as specified on this certificate were tested  
according to the current UL requirements.



Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please  
contact a local UL Customer Service Representative at [www.ul.com/contactus](http://www.ul.com/contactus)

